

Quality Assurance Project Plan

University of Rhode Island Watershed Watch Analytical Laboratory


**Date: V7b February 2024
Prepared for
United States Environmental Protection Agency,
Region 1
New England Laboratory
11 Technology Drive
North Chelmsford, Massachusetts 01863**

**V7 September 2020
V6 April 2018
V5 June 2009,
V4 August 2008,
V3 January 2007,
V2 March 2006,
Original June 2005
Originally prepared for: U.S. Environmental Protection
Agency N.E. Region 1
1 Congress Street, Suite 1100
Boston, Massachusetts 02114-2023**

**Prepared by: University of Rhode Island Watershed
Watch
URI Cooperative Extension Water Quality Program
College of the Environment and Life Sciences
Coastal Institute in Kingston
1 Greenhouse Road
Kingston, Rhode Island, 02881**

Quality Assurance Project Plan

University of Rhode Island Watershed Watch Analytical Laboratory



University of Rhode Island Watershed Watch
URI Cooperative Extension Water Quality Program
College of the Environment and Life Sciences
Coastal Institute in Kingston
1 Greenhouse Road
Kingston, Rhode Island, 02881

Elizabeth M. Herron, Program Director
Linda T Green, Program Director Emeritus
Kelly Addy, Program Coordinator
Arthur J. Gold, Program Advisor Emeritus

Kingston, Rhode Island
2024

Linda Green, M.S., Elizabeth Herron, M.A. Kelly Addy, M.S, and Arthur Gold, Ph.D. are (or were) members of the Dept. of Natural Resources Science, College of the Environment and Life Sciences, University of Rhode Island. **Contribution #5464 of the College of the Environment and Life Sciences**, with support from RI Cooperative Extension, RI Department of Environmental Management, United States Department of Agriculture and local governments. The University of Rhode Island is an equal opportunity employer committed to the principles of affirmative action.

NOTICE OF CHANGES

The following edits/changes have been incorporated to update the previous version (7) of this QAPP:

- Updated authorship, approval and distribution lists
- Updated senior staff and roles
- Added E. coli procedures/tables (SOP 028)
- Updated salinity procedures/tables (SOP 017b)
- Revised equation used for determining MDL
- Expanded data validation section
- Added Laboratory Data Management SOP (027)

For additional information or to request a copy contact:

Elizabeth M. Herron, Program Director
401-874-4552 - eherron@uri.edu

URI Watershed Watch
The Coastal Institute in Kingston
1 Greenhouse Road
Kingston, RI 02881
<https://web.uri.edu/watershedwatch/>

Approval List

URI Watershed Watch Program Director and Laboratory Project Manager – Microbiology

| | |
|---|------|
| Elizabeth Herron University of Rhode Island Watershed Watch Laboratory University of Rhode Island Cooperative Extension 1 Greenhouse Road Kingston, Rhode Island 02881 Phone: 401-874-4552, email: eherron@uri.edu | Date |
|---|------|

URI Watershed Watch Laboratory Project Manager – Nutrients

| | |
|---|------|
| Kelly Addy University of Rhode Island Watershed Watch Laboratory University of Rhode Island Cooperative Extension 1 Greenhouse Road Kingston, Rhode Island 02881 Phone: 401-874-7532, email: Kaddy@uri.edu | Date |
|---|------|

Laboratory Quality Assurance Advisor

| | |
|--|------|
| Dr. Alissa Cox University of Rhode Island Department of Natural Resources Science 1 Greenhouse Road Kingston, Rhode Island 02881 Phone: 401-874-5707, email: alibba@uri.edu | Date |
|--|------|

United States Environmental Protection Agency, Region 1

| | |
|--|--|
| Nora Conlon, Ph.D. United States Environmental Protection Agency New England Laboratory 11 Technology Drive North Chelmsford, Massachusetts 01863 Phone: 617-918-8335 | |
|--|--|

Approval List (continued)

Rhode Island Department of Environmental Management

| | |
|--|------|
| Sue Kiernan – Administrator Rhode Island Department of Environmental Management Office of Water Resources 235 Promenade Street Providence, Rhode Island 02908 Phone: 401-537-4246 | Date |
|--|------|

Rhode Island Department of Health

| | |
|---|------|
| Rajendra Kothavade, Ph.D. Chief Environmental Lab Scientist Division of State Laboratories Rhode Island Department of Health 50 Orms St. Providence, RI 02904 Phone: 401-222-5578 | Date |
|---|------|

Distribution List

University of Rhode Island Watershed Watch Program Director and Laboratory Project Manager – Microbiology

Elizabeth Herron
University of Rhode Island Watershed Watch Laboratory
University of Rhode Island Cooperative Extension
1 Greenhouse Road
Kingston, Rhode Island 02881
Phone: 401-874-4552, email eherron@uri.edu

University of Rhode Island Watershed Watch Laboratory Project Manager – Nutrients

Kelly Addy
University of Rhode Island Watershed Watch Laboratory
University of Rhode Island Cooperative Extension
1 Greenhouse Road
Kingston, Rhode Island 02881
Phone: 401-874-7532, email: Kaddy@uri.edu

Laboratory Quality Assurance Advisor

Alissa Cox
University of Rhode Island
Department of Natural Resources Science
1 Greenhouse Road
Kingston, Rhode Island 02881
Phone: 401-874-5707, alibba@uri.edu

Cooperating Agencies and Organizations:

United States Environmental Protection Agency, Region 1

Nora Conlon, Ph.D.
United States Environmental Protection Agency
New England Laboratory
11 Technology Drive
North Chelmsford, Massachusetts 01863
Phone: 617-918-8335

Rhode Island Department of Environmental Management

Sue Kiernan – Administrator
Office of Water Resources
235 Promenade Street
Providence, Rhode Island 02908
Phone: 401-537-4160

Distribution List (continued)

Rhode Island Department of Environmental Management

Jane Sawyers – Water Quality Standards
Office of Water Resources
235 Promenade Street
Providence, Rhode Island 02908
Phone: 401-222-4700 ext. 7300

Rhode Island Department of Health

Rajendra Kothavade, Ph.D.
Chief Environmental Lab Scientist
Division of State Laboratories
Rhode Island Department of Health
50 Orms Street
Providence, RI 02904
Phone: 401-222-5578

University of Rhode Island Coastal Institute

Nathan Vinhateiro - Assistant Director, Science & Data Management
Room 101, 1 Greenhouse Road
Kingston, Rhode Island 02881
Phone: 401-874-4794

Wood-Pawcatuck Watershed Association

Christopher Fox – Executive Director
203B Arcadia Rd.
Hope Valley, Rhode Island 02852

Narrow River Preservation Association

Annette DeSilva – Monitoring Coordinator
56 South River Road
Narragansett, Rhode Island 02882

Salt Ponds Coalition

Ms. Alicia Eichinger
Executive Director
PO Box 375
Charlestown, Rhode Island 02813

Save Bristol Harbor

Barbara Healy, Monitoring Coordinator
PO Box 242
Bristol, RI 02809

Distribution List (continued)

Clean Up the Sound and Harbors

Jack Leary, Monitoring Coordinator
PO Box 883
Stonington, CT 06378

Rhode Island Rivers Council

Ms. Alicia Eichinger, chair
c/o RI Water Resources Board
RI Department of Administration
One Capitol Hill
Providence, RI 02908

Narragansett Indian Tribe

Dinalyn Spears
Director of Community Planning/Natural Resources Dept.
P.O. Box 268
Charlestown, RI 02813

List of Abbreviations

| Abbreviation | Definition |
|----------------|---|
| CA | Corrective Action |
| COC | Chain-of-Custody |
| %D | Percent Difference |
| DI | Deionized Water |
| DQIs | Data Quality Indicators |
| DO | Dissolved Oxygen |
| DQO | Data Quality Objectives |
| EPA-NE | Environmental Protection Agency – New England District (Region 1) |
| g | Gram |
| L | Liter |
| LCS | Laboratory Control Standard (standard analyzed as a sample) |
| MDL | Method Detection Limit |
| mL | Milli-liter |
| mg | Milli-gram |
| MSDS | Material Safety Data Sheet |
| MQ water | Ultra-pure water >18uohms |
| NA | Not Applicable |
| QAPP | Quality Assurance Project Plan |
| QA/QC | Quality Assurance/Quality Control |
| ppb | Part per billion (ug/L) |
| ppm | Part per million (mg/L) |
| R ² | Coefficient of Determination |

List of Abbreviations (continued)

| Abbreviation | Definition |
|--------------|--|
| %RPD | Replicate Percent Difference |
| RL | Reporting Limit (Quantitation Limit) |
| SOP | Standard Operating Procedure |
| SU | Standard Unit (pH units) |
| µg | Micro-gram |
| URIWW | University of Rhode Island Watershed Watch |

TABLE OF CONTENTS

| | | |
|-----------------|---|-----------|
| 1.0 | PURPOSE AND DESCRIPTION | 1 |
| 1.1 | QUALITY ASSURANCE PROJECT PLAN (QAPP) OBJECTIVES | 4 |
| 1.2 | ORGANIZATION AND COMMUNICATION | 4 |
| 1.2.1 | <i>Staff Responsibilities</i> | 5 |
| 1.2.1.1 | Program Director URIWW Program and Laboratory Manager | 5 |
| 1.2.1.2 | Program Coordinator | 6 |
| 1.2.1.3 | Laboratory Manager - Nutrients | 6 |
| 1.2.1.4 | Laboratory Manager – Microbiology | 6 |
| 1.2.1.5 | QA Advisor | 7 |
| 1.2.1.6 | Student Laboratory Technicians | 7 |
| 1.2.2 | <i>Personnel Qualifications</i> | 8 |
| 1.2.1 | <i>Training</i> | 8 |
| 1.3 | SCHEDULE/TIMELINE | 10 |
| 2.0 | LABORATORY QUALITY OBJECTIVES AND MEASUREMENT PERFORMANCE | |
| CRITERIA | 10 | |
| 2.1 | METHOD DETECTION LIMITS (MDL) AND REPORTING LIMIT (RL) | 10 |
| 2.1.1 | <i>Method Detection Limit (MDL) calculation</i> | 10 |
| 2.2 | PRECISION | 12 |
| 2.3 | ACCURACY | 12 |
| 2.4 | COMPARABILITY | 12 |
| 2.5 | COMPLETENESS | 13 |
| 2.6 | QA/QC TABLES | 13 |
| 2.6.1 | <i>Contaminants of Concern and Other Target Analytes Table (Reference Limit and Evaluation Table) - Worksheet #9b</i> | 14 |
| 2.6.2 | <i>Measurement Performance Criteria Table – Worksheet 11b</i> | 16 |
| 2.6.3 | <i>Fixed Laboratory Analytical Method/SOP Reference Table – Worksheet 20</i> | 23 |
| 2.6.4 | <i>Fixed Laboratory Instrument Maintenance and Calibration Table - Worksheet 21</i> | 25 |
| | <i>Note: Maintenance and service record forms are in Appendix D</i> | 32 |
| | <i>Fixed Laboratory Analytical QC Sample Table –Worksheet #24 a</i> | 33 |
| 3.0 | SAMPLE HANDLING, TRACKING AND CUSTODY REQUIREMENTS | 44 |
| 3.1 | LABORATORY SAMPLE TRACKING | 45 |
| 3.2 | ACCEPTANCE OF EXPENDABLE LABORATORY SUPPLIES | 46 |
| 3.3 | SAMPLE HANDLING SYSTEM – WORKSHEET 16 | 47 |
| 4.0 | PROJECT DOCUMENTATION AND RECORDS | 48 |
| 5.0 | DATA VALIDATION | 48 |

LIST OF FIGURES
(Tables listed in Table of Contents)

| | |
|---|----|
| Figure 1 - University of Rhode Island Watershed Watch Laboratory Structure..... | 5 |
| Figure 2 – Student Training Record..... | 9 |
| Figure 3 – Sample Log Sheet | 44 |
| Figure 4 - Internal Laboratory Tracking Scheme..... | 45 |

LIST OF APPENDICIES

| | |
|---|--|
| Appendix A – Standard Operating Procedures | |
| Appendix B – Resumes of Key Laboratory Personnel | |
| Appendix C – Descriptive Information Regarding University of Rhode Island Watershed Watch | |
| Appendix D – Maintenance, Service and Training Record Forms | |
| Appendix E – Laboratory Floor Plans | |
| Appendix F – URIWW List of Major Laboratory Equipment | |

1.0 PURPOSE AND DESCRIPTION

The University of Rhode Island Watershed Watch Program (URIWW) is a Cooperative Extension Water Quality Program in the Department of Natural Resources Science, College of the Environment and Life Sciences. The program is located in the Coastal Institute building on the URI Kingston campus. Begun in 1988, the URIWW program is a statewide volunteer monitoring program with over 300 volunteers. The program focuses on providing current information on the water quality of surface water resources throughout Rhode Island. It is a service provider to statewide and local decision-makers and is the sole source of long-term lake water quality data for Rhode Island. The URIWW laboratory provides analytical services to the Rhode Island Department of Environmental Management (RIDEM) and the Environmental Protection Agency, New England District (Region 1) (EPA-NE) as well as other URI researchers. It is a springboard for municipal board activities by volunteers, linked with all cooperative extension water quality activities. The program is intended to encourage communities and shoreline residents to understand the need to cooperatively manage and improve the water quality of all the water bodies within a watershed.

Information describing the URIWW program, program factsheets, water quality data as well as monitoring protocols are maintained at the following web-site:
<https://web.uri.edu/watershedwatch/>. Water quality data are available in CSV format suitable for downloading and use by water resource professionals. Data are also provided in a variety of formats designed to be more accessible to a lay audience, including an interactive data dashboards, tables and charts by site or parameter, and watershed maps. Basic information describing URIWW is also available in Appendix C.

The purpose of this Quality Assurance Project Plan (QAPP) is to provide guidance on the analytical procedures and quality assurance/quality control (QA/QC) tasks performed by the URIWW. The URIWW Laboratory provides analysis of samples for the following contaminants of concern: fecal coliform, enterococci, E. coli, salinity, alkalinity, pH, chlorophyll-a, chloride, ammonia, orthophosphate, nitrate + nitrite, total phosphorus and total nitrogen. Assays are completed on ambient waters (rivers, lakes and streams) and marine waters (ocean, estuaries, brackish).

This QAPP does not describe field collection nor analysis procedures; that information is provided in other documents. A cross-reference between the information required by EPA-NE is provided in the table below. Note that information found in narrative format instead of in an EPA-NE table is listed as “in narrative”.

Required Information Checklist

| EPA-NE Work- sheet number | Worksheet Title | Location In URIWW Laboratory QAPP |
|--|---|--|
| 1 | Title and approval | In narrative |
| 2 | Table of contents & document format | In narrative |
| 3 | Distribution list | In narrative |
| 4 | Project personnel sign-off sheet | All relevant personnel are included on the approval page |
| 5a | Organizational chart | Figure 1 |
| 5b | Communication pathway | Section 1.2 in narrative |
| 6 | Personnel responsibilities and qualification | Section 1.2 and 1.2.1 in narrative |
| 7 | Special personnel training requirements | Section 1.2.2 in narrative |
| 8a | Project scoping meeting attendance sheet, agenda | NA |
| 8b | Problem definition/site history & background | Section 1.0 in narrative |
| 9a | Project description | Section 1.0 in narrative |
| 9b | Contaminants of concern | Section 2.6.1 |
| 9c | Field & QC sample summary | NA |
| 10* | Project schedule timeline | Section 1.3 in narrative |
| 11a | Project quality objectives/decision statements | Section 2.0 in narrative |
| 11b | Measurement performance criteria table | Section 2.6.2 |
| 12a | Sampling design & rationale | NA |
| 12b | Sampling locations, methods, SOP requirements table | NA |

| EPA-NE Work- sheet number | Worksheet Title | Location In URIWW Laboratory QAPP |
|--|---|--|
| 13 | Project sampling SOP table | Appendix A |
| 14 | Field equipment calibration | NA |
| 15 | Field equipment maintenance | NA |
| 16 | Sampling handling, tracking, custody | Section 3.0 in narrative and Section 3.2 |
| 17 | Field method /SOP | NA |
| 18 | Field calibration | NA |
| 19 | Field maintenance | NA |
| 20 | Fixed lab. analytical , SOP reference table | Section 2.6.3 |
| 21 | Lab instrument maintenance & calibration table | Section 2.6.4 |
| 22a | Field sampling QC | NA |
| 22b | Field sampling QC continued | NA |
| 23a | Field analytical QC | NA |
| 23b | More field QC | NA |
| 24a | Lab analytical QC | Section 2.6.5 |
| 24b | More lab analytical QC | No multiple analytes |
| 25 | Non-direct measurement criteria | NA |
| 26 | Project documentation and records | Section 4.0 in narrative |
| 27a | Assessment and response | NA |
| 27b | Project assessment | NA |
| 27c | Project assessment plan | NA |
| 28 | QA management reports | Section 4.0 in narrative |
| 29a | Data evaluation process | NA |

| EPA-NE Work- sheet number | Worksheet Title | Location In URIWW Laboratory QAPP |
|--|-------------------------------|--|
| 29b | Data validation summary | Section 5.0 in narrative |
| 29c | Data validation modifications | NA |
| 30 | Data usability assessment | NA |

Notes:

NA – Not applicable to this QAPP. This QAPP provides information regarding general laboratory protocols only. No project-specific information is contained in this general QAPP. No field sample collection or analysis information is provided in this QAPP and all data are generated in-house.

1.1 Quality Assurance Project Plan (QAPP) Objectives


The objective of this QAPP is to present the organization, objectives and specific quality assurance/quality control (QA/QC) procedures associated with URIWW laboratory analysis protocols. Guidance on the analysis procedures for the following laboratory assays is provided in this document: fecal coliform, E. coli, enterococci, alkalinity, pH, salinity, chlorophyll-a, chloride, ammonia, orthophosphate, nitrate + nitrite nitrogen, total phosphorus and total nitrogen. Specific QA/QC criteria as well as documentation are outlined in individual Standard Operation Procedures (SOPs) located in Appendix A. This QAPP does not describe any field collection or analysis procedures; this information is provided in other documents.

1.2 Organization and Communication

Elizabeth Herron is the URIWW Program Director as well as the overall Analytical Laboratory manager, and Laboratory Manager for microbiological analysis (Figure 1). She is responsible for the analysis and QA/QC of microbiological assays. She is also responsible for overall operation of the laboratory, including QA/QC of all non-nutrient and chloride related assays, and data management for all laboratory analyses and field data. Kelly Addy is the Watershed Hydrology Laboratory Manager and URIWW Program Coordinator and Nutrient Analytical Laboratory manager. She is responsible for overall management of the nutrient analyses, including QA/QC of all nutrient and chloride assays. Dr. Alissa Cox will provide QA/QC guidance as the URIWW QA Advisor. Ms. Herron and Ms. Addy are both responsible for the supervision of student laboratory technicians.

All changes to the QAPP or specific SOPs will be completed only after review and acceptance by Ms. Herron or Ms. Addy.

Figure 1: University of Rhode Island Watershed Watch (URIWW) Laboratory Structure



1.2.1 Staff Responsibilities

The responsibilities of each staff position are described below.

1.2.1.1 Program Director URIWW Program and Laboratory Manager

Responsibilities and duties include:

- Daily operations and periodic review of URIWW program;
- Obtaining and managing grants and contracts;
- Updating of technical documentation when necessary including: establishment of acceptance criteria, monitoring corrective action, oversight of the QA/QC program and arranging for annual internal audits of operation;
- Being present on laboratory premises during most laboratory hours of operation to ensure adequate and appropriate supervision of laboratory activities;
- Ensuring accurate performance of all tests in the laboratory including submission of appropriate reports;
- Ensuring adequate supervision of laboratory staff and the hiring of adequately trained personnel;
- Adhering to written QA/QC procedures;
- Maintaining employee records;
- Maintaining compliance with relevant URI Health and Safety requirements.
- Coordination of volunteer monitors including equipment, supplies and sample delivery;
- In charge of data templates, data management, database implementation and data entry;

Qualifications:

- Masters degree in chemical, biological or environmental sciences;
- A minimum of four (4) years of analytical laboratory experience.
- Extensive experience working with volunteers on water quality projects.

1.2.1.2 Program Coordinator

Responsibilities and duties include:

- Updating technical documentation when necessary;
- Creating outreach tools and data reports;
- Assisting with obtaining and managing grants and contracts;
- Developing and maintaining program website;
- Ensuring adequate supervision of laboratory staff;
- Completing duties of the Program Director when they are absent from the laboratory;
- Assisting the program director with assigned tasks.

Qualifications:

- Masters degree in chemical, biological or environmental sciences;
- A minimum of two (2) years of analytical laboratory experience;
- Extensive experience working with volunteers on water quality projects.

1.2.1.3 Laboratory Manager - Nutrients

Responsibilities and duties include:

- Supervising and/or completing preparation and analysis of nutrients and chloride samples;
- Review of all nutrient data;
- Assuring that proper QA/QC procedures are followed as documented in laboratory QAPP;
- Maintaining proper instrument maintenance records;
- Documenting laboratory procedures;
- Ordering laboratory supplies;
- Maintaining compliance with URI Health and Safety requirements;
- Attending initial URI Health and Safety training: Environmental Awareness/Initial Laboratory Waste Management Training and the yearly refresher: Prudent Practices and Laboratory Waste Management.

Qualifications:

- Bachelors degree in chemical, biological or environmental sciences;
- Minimum of two (2) years of analytical laboratory experience.

1.2.1.4 Laboratory Manager – Microbiology

Responsibilities and duties include:

- Supervising and/or completing preparation and analysis of all microbiological samples;
- Supervising and/or completing preparation and analysis of all samples other than nutrient samples (including pH, alkalinity, chlorophyll, salinity, dissolved oxygen, and preparation of nutrient samples);
- Review of all program data;

- Training interns and student employees as well as documenting training activities;
- Assuring that proper QA/QC procedures are followed as documented in the laboratory QAPP;
- Maintaining proper instrument maintenance records;
- Documenting laboratory procedures;
- Ordering laboratory supplies;
- Maintaining compliance with URI Health and Safety requirements;
- Attending the initial URI Health and Safety training: Environmental Awareness/Initial Laboratory Waste Management Training and the yearly refresher: Prudent Practices and Laboratory Waste Management.

Qualifications:

- Bachelors degree in chemical, biological or environmental sciences;
- Minimum of two (2) years of analytical laboratory experience.

1.2.1.5 QA Advisor

Responsibilities and duties include:

- Providing technical assistance in determining appropriate QA/QC measures when requested by program director or coordinator.

Qualifications:

- Masters degree in chemical, biological or environmental sciences;
- Minimum of two (5) years of related experience;
- Extensive knowledge of QA/QC procedures.

1.2.1.6 Student Laboratory Technicians

Responsibilities and duties include:

- Labware cleaning and preparation for use;
- Organizing, set up and distribution of samples bottles;
- Logging in samples
- Cleaning, calibrating, break down and set up of monitoring supplies;
- Preparation and analysis of samples for pH, alkalinity, salinity and chlorophyll-a, and including occasional microbiological samples;
- Data entry and proof-reading data entries
- Properly following established QA/QC procedures;
- Following proper URI Health and Safety requirements;
- Maintaining compliance with URI Health and Safety requirements by attending initial student Health and Safety training.

Qualifications:

- High school diploma
- Either a bachelor's degree in chemical, biological or environmental science or be currently working toward a bachelors in chemical, biological or environmental science.

1.2.2 Personnel Qualifications

A brief description of the experience of principal laboratory personnel is described here. Resumes of key personnel are in Appendix B.

Elizabeth Herron has more than 30 years of experience with the URIWW program, beginning by completing a QA assessment of the program to determine the representativeness of the data generated. She expanded the microbiology laboratory, updating it to current standards, enabling it to receive state-certification. She is the recipient of numerous grants and awards related to her work with URIWW. She is a former director of the North American Lake Management Society as well as a co-founding member of the Rhode Island Volunteer Monitoring Steering Committee. She has authored numerous articles and technical publications and has presented workshops, technical papers and webinars throughout the United States.

Kelly Addy is a watershed hydrologist who has over managed the Watershed Hydrology Laboratory (WHL) for over 20 years. She focuses on watershed sources and sinks of nitrogen, completing nutrient analyses for thousands of samples over the decades. She has authored many refereed articles and received numerous grants and awards for her work with the WHL, including the prestigious URI College of the Environment and Life Sciences Research Staff Research and Scholarship Excellence Award in 2019. Ms. Addy has assisted URIWW throughout her career and is well associated with its practices.


Dr. Alissa Cox is a Clinical Assistant Professor of Sustainable Ecological Design, and director of the New England Onsite Wastewater Training Program in the College of the Environment and Life Sciences at URI. She has more than a decade of experience in the field of water resources, particularly focused on the impact of on-site wastewater treatment systems to Rhode Island waters. She has published numerous refereed journal articles related to her wastewater work. In addition, she has an extensive background in science and math education, which is especially valuable for assessing training procedures. As head of the Laboratory of Soil Ecology and Microbiology, she is experienced and familiar with the procedures and instruments used by URIWW.

Linda Green has over 35 years of analytical laboratory related experience and was the director of URIWW for more than 30 years from its inception in 1988. She is the recipient of numerous awards and grants related to her work with the URIWW program and has authored numerous articles and technical publications. Ms. Green has hosted workshops on QA/QC in volunteer monitoring programs and for ten years was the sole volunteer monitoring representative on the National Water Quality Monitoring Council as well as a co-founding member of the Rhode Island Volunteer Monitoring Steering Committee. She continues to consult with URIWW ensuring that this vital and extensive institutional knowledge is not lost.

1.2.1 Training

Except for the mandatory URI safety training, laboratory personnel training is conducted by Elizabeth Herron. Laboratory training is provided on basic laboratory techniques as well as method-specific details. Training requirements for each assay are provided in analyte-specific SOPs, located in Appendix A. Detailed written procedures are posted in the vicinity of all procedure spaces to provide additional post-training support. All laboratory assays are conducted by laboratory personnel; no volunteer monitors conduct laboratory assays.

Figure 2 – Student Training Record



**URI Watershed Watch Laboratory
Student Training Record**

Student Name : _____

| Method | Date of Training | Initials of Trainer | Initials of Student |
|--|------------------|---------------------|---------------------|
| Field | | | |
| Attended Watershed Watch Volunteer training | | | |
| General | | | |
| SOP 001 - URI General Laboratory Safety SOP 001a - University Safety & Waste Handling | | | |
| SOP 002 - Laboratory Water SOP 003 - General Labware Cleaning Procedure | | | |
| SOP 004 - General Autoclave Operation SOP 005 - Bottle Autoclaving Procedure SOP 006 - Waste Autoclaving Procedure | | | |
| SOP 019 - Analytical Balance Calibration | | | |
| SOP 025 - Lab Thermometer Calibration | | | |
| Chemistry and Particulates | | | |
| SOP 012 - Chlorophyll - a Analysis, Trilogy | | | |
| Sample preparation for SOP 016 - Total Phosphorus and Nitrogen Analysis | | | |
| SOP 017b - Salinity Analysis using Digital Refractometer | | | |
| SOP 021 - pH Procedures | | | |
| SOP 022 - Alkalinity Procedures | | | |
| SOP 023 - Filtering Water Samples | | | |
| Microbiology | | | |
| SOP 018 - Enterococci Analysis - Enterolert | | | |
| SOP 024 - Fecal coliform Analysis Coli-18 | | | |
| SOP 026 - HPC Quanti-tray Analysis | | | |
| SOP 028 - E. coli Analysis Coli-18 | | | |

This form is intended to be used as a training record. This record will be maintained with student employment records by URI Watershed Watch for at least 5 years.

Training of laboratory personnel is recorded on the “Student Training Record” document, figure 2 above. Records of training are archived for at least 5 years.

1.3 Schedule/Timeline

This QAPP does not relate to a specific project, therefore no specific timeline or schedule is offered. Laboratory analyses are conducted as needed to provide information for partners, sponsors and various projects.

2.0 LABORATORY QUALITY OBJECTIVES AND MEASUREMENT PERFORMANCE CRITERIA

High quality data is the goal of all URIWW Laboratory analyses. Specific data quality objectives have been set on a method basis for method detection limits (MDL), precision, accuracy, comparability and completeness. Values specific to each of these objectives are located in analyte-specific SOPs located in Appendix A as well as below. Since this document is a general QAPP for laboratory assays only, there are no specific if/then statements linking laboratory criteria to project decisions.

2.1 Method Detection Limits (MDL) and Reporting Limit (RL)

The MDL is the analyte concentration where there is 99% confidence that the sample concentration is different than zero. Below the MDL it is uncertain if the concentration is not zero. The reporting limit (RL) is the value above which data have definable accuracy and precision. Each analyte of interest has a specific MDL and RL value. These values are located in the analyte-specific SOPs in Appendix A as well as worksheet 9b (see Section 2.6.1). Note that most MDL's and RL's are calculated annually, the most up-to-date values are available from the Laboratory Manager.

The analytical method MDL as reported in Section 2.6.1 for each assay is often different from the achievable laboratory MDL. Generally, the achievable laboratory MDL is higher than the analytical method MDL. This is often the case because the MDL listed for an analytical method is for the best-case scenario. In this scenario, there are no other contaminants present in a sample that could cause interferences during sample analysis, the method blank would be extremely low and all equipment would function without error. Unfortunately, this is generally not the case. At the very low contaminant levels that the laboratory is able to analyze samples to it is easy to introduce some contamination from water or reagents. Therefore, the URIWW laboratory elevates the method MDL to the RL to account for these concerns.

2.1.1 Method Detection Limit (MDL) calculation

The MDL is calculated for the following assays annually: chloride, ammonia, orthophosphate, nitrate+nitrite nitrogen, total nitrogen and total phosphorus. The MDL is calculated using the method found in 40 CFR part 136 Appendix B (Appendix B to Part 136 – Definition and Procedure for the Determination of the Method Detection Limit – Revision 1.11). Calculation of the MDL for the microbiology assays is dependent upon the amount of sample filtered or processed and is discussed in SOP 018 - Enterococci Using Enterolert IDEXX Method and SOP-024 – Fecal Coliform Analysis Using Colilert 18 (IDEXX method). The method detection limits for alkalinity, pH, chlorophyll-a and salinity are set by the method or the associated instrument and are not recalculated yearly.

The following provides a brief description of the method used to calculate MDL from 40 CFR part 136 Appendix B.

1. The MDL is estimated as the concentration 3 times the standard deviation of replicate measurements of a standard prepared in reagent water.
 - a. The standard deviation is calculated as follows:

$$S^2 = \frac{1}{n-1} \left[\sum_{i=1}^n x_i^2 - \frac{\left(\sum_{i=1}^n x_i \right)^2}{n} \right]$$

$$s = (S^2)^{\frac{1}{2}}$$

Where:

S^2 = Variance

S = Standard deviation

N = Number of samples

X_i = Sample value $i = 1$ to n , are the analytical results in the final method reporting units obtained from the n sample aliquots and \sum refers to the sum of the $i = 1$ to n

2. Prepare a standard in reagent water that is 1 to 5 times the concentration of the estimated MDL (which is 3 to 5 times the standard deviation as calculated above).
3. Analyze a minimum of 7 aliquots of this standard. If a blank measurement is required to determine the final sample concentration, analyze one blank per aliquot. Then use the average concentration of the blank in the final calculation to determine sample concentration.
4. Calculate the standard deviation of the final concentration of the seven aliquots from step 3.
5. Compute the MDL as follows:

$$a. \text{ MDL} = t_{(n-1, 1-\alpha = 0.99)} \times s$$

Where: s = Standard deviation

$t_{(n-1, 1-\alpha = 0.99)}$ = Students' t value appropriate for the 99% confidence level and a standard deviation estimate with $n-1$ degrees of freedom

Table 1 - Students' T Values At The 99% Confidence Level

| Number of replicates | Degrees of freedom (n-1) | $t_{(n-1, .99)}$ | Number of replicates | Degrees of freedom (n-1) | Number of replicates |
|----------------------|--------------------------|------------------|----------------------|--------------------------|----------------------|
| 7 | 6 | 3.143 | 10 | 9 | 2.821 |
| 8 | 7 | 2.998 | 11 | 10 | 2.764 |
| 9 | 8 | 2.896 | 16 | 15 | 2.602 |

6. If the calculated MDL is higher than expected, repeat the procedure with a standard of lower concentration.
7. The final MDL should be 1/3rd to 1/5th the calibrator used, or the lowest calibrator should be 3 to 5 times the estimated MDL.

2.2 Precision

Precision is an evaluation of the degree to which two or more measurements are in agreement as well as a measurement of random error. Precision will be assessed through the measurement of duplicate samples and subsequent calculation of the relative percent difference (%RPD) as described below.

$$\%RPD = \frac{|\text{Result of Replicate 1} - \text{Result of Replicate 2}|}{\text{Average of Result of Replicate 1 and Result of Replicate 2}} \times 100$$

Objectives for precision are located in the analyte specific SOPs (Appendix A) as well as worksheet 11b and 24a Section 2.6.2 and 2.6.5, respectively.

2.3 Accuracy

Accuracy is an evaluation of the degree to which a measured value and a known reference value or true value are in agreement. This is a measurement of systematic error and is often referred to as “bias”. Laboratory accuracy is determined by the analysis of reference material and comparison of the resulting value to that of the accepted value. The difference between the accepted and reference value is the percent difference (%D). The %D is calculated as follows:

$$\%D = \frac{|\text{Known Value of Reference Material} - \text{Calculated Value of Reference Material}|}{\text{Known Value of Reference Material}} \times 100$$

Objectives for accuracy are located in the analyte specific SOPs (Appendix A) as well as worksheet 11b and 24a (Section 2.6.2 and 2.6.5, respectively).

Accuracy is determined during both routine sample analysis procedures as well as by yearly participation in the EPA Water Pollution Proficiency Test Study for the following assays: alkalinity, ammonia, chloride, nitrate + nitrite-N, orthophosphate, total nitrogen, total phosphorus, and pH.

2.4 Comparability

All methods utilized by the URIWW Laboratory are based on methods found in *Standard Methods for the Examination of Water and Wastewater* published by the American Public Health Association, American Water Works Association and Water Environment Federation. Specific references for each method are found in the analyte specific SOPs (Appendix A).

2.5 Completeness

Completeness is a measure of the amount of valid data obtained from the laboratory methods compared to the amount that was expected to be obtained under normal conditions. Greater than 90% completeness of accepted field samples is expected. Completeness is calculated as follows:

$$\text{Completeness} = \frac{\text{Number of Valid Laboratory Measurements}}{\text{Number of Laboratory Measurements Planned}} \times 100$$

2.6 QA/QC Tables

Tables summarizing the QA/QC objectives for each analysis performed by the URIWW Laboratory are provided on the following pages. These tables specifically address the Data Quality Indicators (DQIs) or the procedures to be followed to provide assurance that an analytical procedure is returning valid results. Each DQI has a specific result that must be met before the data is considered acceptable. Information is also provided on the instruments utilized for each assay and the maintenance and calibration procedures that must be completed on each instrument. Analyte-specific tables provide information on the number of QA/QC samples to be prepared (blanks, replicates, etc.) and the expected result as well as the person(s) responsible for assessing any problems and determining the proper course of action, if necessary.

2.6.1 Contaminants of Concern and Other Target Analytes Table (Reference Limit and Evaluation Table) - Worksheet #9b

| EPA-NE QAPP Worksheet #9b – EPA Rev. 10/99 Contaminants of Concern and Other Target Analytes Table (Reference Limit and Evaluation Table) | | | | | | | | |
|---|------------|--|---|---|-------------------|------------|------------------------------|------------------|
| Analyte | CAS Number | Reporting Units | Project Action Limit (Units) (wet or dry weight) | Project Quantitation Limit (Units) (wet or dry weight) | Analytical Method | | Achievable Laboratory Limits | |
| | | | | | MDLs | Method RLs | MDLs ³ | RLs ³ |
| Fecal coliforms – SOP 024 | | number/100mL | | | <1 ¹ | | <1 ¹ | |
| Enterococci – SOP 018 | | number/100mL | | | <1 ¹ | | <1 ¹ | <1 ¹ |
| Heterotrophic Plate Count – SOP 020 | | number/100mL | | | <1 ² | | <1 ² | <1 ² |
| Alkalinity – SOP 010 | | mg/L CaCO ₃ | | | Not Provided | | 0 | 0.1 |
| pH – SOP 010 | | Standard Unit (SU) | | | 1.0 | | 1.00 | 1.0 |
| Salinity – SOP 017b | | ppt | | | Not Provided | | 1.0 | 1.0 |
| Chlorophyll a – SOP 012 | | µg/L chlorophyll-a | | | 0.1 | | 0.1 | 0.2 |
| Chloride – SOP 013 | 16887-00-6 | mg/L Cl ⁻ | | | 0.2 | | 0.13 | 1 |
| Ammonia – SOP 014 | 7664-41-7 | µg/L NH ₃ -N | | | 5 | | 3.0 | 20 |
| Orthophosphate – SOP 015 | | µg/L PO ₄ -P | | | 2 | | 0.4 | 4 |
| Nitrite + Nitrate – SOP 015 | | µg/L NO ₃ /NO ₂ -N | | | 10 | | 1.5 | 15 |

EPA-NE QAPP Worksheet #9b – EPA Rev. 10/99

Contaminants of Concern and Other Target Analytes Table (Reference Limit and Evaluation Table)

| Analyte | CAS Number | Reporting Units | Project Action Limit (Units) (wet or dry weight) | Project Quantitation Limit (Units) (wet or dry weight) | Analytical Method | | Achievable Laboratory Limits | |
|----------------------------|------------|-----------------|--|--|-------------------|------------|------------------------------|------------------|
| | | | | | MDLs | Method RLs | MDLs ³ | RLs ³ |
| Total Phosphorus – SOP 016 | | µg/L P | | | 2* | | 0.4 | 4 |
| Total Nitrogen – SOP 016 | | µg/L N | | | 10* | | 2 | 20 |

Notes:

*The MDLs for Total Phosphorus and Total Nitrogen were not provided by the method reference. Therefore, the method MDLs for orthophosphate and nitrate + nitrite were reported since the total phosphorus and nitrogen assays are based on the orthophosphate and nitrate + nitrite assays, respectively. After a sample is digested for total nitrogen and phosphorus the sample is analyzed as a nitrate + nitrite and orthophosphate sample (please refer to the analyte-specific SOP located in the URIWW Laboratory QAPP for more information).

¹Taken from MPN Tables for 51 Well Quanti-Tray and Quanti-Tray/2000 - www.IDEXX.com accessed December, 2006

²Taken from HPC for Quanti-tray table <https://www.idexx.com/en/water/water-products-services/hpc-quant-tray/> accessed September 24, 2020

³Some MDLs and RLs are determined annually, for the most up-to-date values contact the Laboratory Manager

2.6.2 Measurement Performance Criteria Table – Worksheet 11b

Note: All QC Measurement Performance Criteria in this table are for assessment of analytical error only.

| EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99 Measurement Performance Criteria Table | | | | |
|---|---|---|--------------------------------|-----------------------|
| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
| Fecal coliform, Enterococci | Method Blank | < 1 /100 mL | Bias | 024/018 |
| | Sample Replication | Not greater than 20%RPD | Precision | |
| | Inoculate a tray with a known positive sample (method for positive trays and QA check on new reagent batches) | Positive growth | Bias/false negatives | |
| | Check of UV sterilizer efficiency | Not less than 70% of initial efficiency | False positives | |
| | Sample bottle sterility checks | 0 \number/mL | Bias/false positives | |
| | Check incubator temperature | 41 °C Incubator 41 +/- 0.5 °C 35 °C Incubator 35 +/- 0.5 °C 44.5 °C Incubator 44.5 +/- 0.2 °C | Bias | |
| | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/Comparability | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|---|---|--|--------------------------------|-----------------------|
| Fecal coliform, Enterococci (continued) | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | 024/018 |
| | IDEXX Mixing Bottle Sterility Check | 0 number/mL | Bias/False positives | |
| Heterotrophic Plate Count (HPC) | Method Blank | < 1 colony/mL | Bias | 020 |
| | Check of UV sterilizer efficiency | Not less than 70% of initial efficiency | False positives | |
| | Sample bottle sterility checks | 0 number/mL | Bias/false positives | |
| | Check incubator temperature | 35 °C Incubator 35 +/- 0.5 °C | Bias | |
| Alkalinity and pH | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | 010 |
| | Calibration | Electrode efficiency greater than 96% | Accuracy | |
| | Sample Replication | pH – difference not greater than +/- 0.5 S.U. Alkalinity – difference not greater than 25%D | Precision | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|-------------------------------|--|---|-----------------------------------|--------------------------|
| Alkalinity and pH (continued) | Standards as Samples (Calibration check) | Change in 7.0 standard not greater than +/- 0.2 SU | Accuracy/ Precision | 010 |
| Salinity | Sample Replication | Not greater than 2 ppt different | Precision | 017b |
| | Sample Comparison | Not greater than 2 ppt different | Accuracy/ Comparability | |
| Chlorophyll-a | Method Blank | Not greater than 0.03 µg/L chlorophyll-a as read on the fluorometer | Bias | 012 |
| | Filter Blank | Not greater than 0.03 µg/L chlorophyll-a as read on the fluorometer | Bias | |
| | Initial Calibration Using Liquid Standards | N/A | Accuracy | |
| | LCS (Calibration check using Solid Standard) | Not greater than 15%D | Accuracy/ Precision | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|--------------------|---|--|-----------------------------------|--------------------------|
| Chloride | Method Blank | Not greater than 2 mg/L Cl ⁻ | Bias | 013 |
| | Sample Replication | Not greater than 20%RPD (Replicate from same sample cup) Not greater than 20%RPD (Replicate from different sample cups) | Precision | |
| | Calibration | R ² of calibration linear regression not less than 0.990 | Accuracy | |
| | Matrix Spike | 80 – 120 % recovery | Bias | |
| | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | |
| | Laboratory Control Samples (Purchased External Standards) | Not greater than 20%D | Accuracy/ Comparability | |
| | Standards as Samples (Calibration check) | Not greater than 20%D | Accuracy | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|--------------------|---|--|-----------------------------------|--------------------------|
| Ammonia | Method Blank | Not greater than 30 µg/L NH ₃ -N | Bias | 014 |
| | Sample Replication | Not greater than 20%RPD (Replicate from different sample cups) | Precision | |
| | Calibration | R ² of calibration linear regression not less than 0.990 | Accuracy | |
| | Matrix Spike | 80 – 120 % recovery | Bias | |
| | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | |
| | Laboratory Control Samples (Purchased External Standards) | Not greater than 20%D | Accuracy/ Comparability | |
| | Standards as Samples (Calibration check) | Not greater than 20%D | Accuracy | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|------------------------------------|---|--|--------------------------------|-----------------------|
| Orthophosphate & Nitrate + Nitrite | Method Blanks | Not greater than 4 µg/L PO ₄ -P and 20 µg/L NO ₃ /NO ₂ -N | Bias | 015 |
| | Sample Replication | Not greater than 20%RPD (Replicate from different sample cups) | Precision | |
| | Matrix Spike | 80 – 120 % recovery | Bias | |
| | Calibration | R ² of calibration linear regression not less than 0.990 | Accuracy | |
| | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | |
| | Laboratory Control Samples (Purchased External Standards) | Not greater than 20%D | Accuracy/ Comparability | |
| | Standards as Samples (Calibration check) | Not greater than 20%D | Accuracy | |

EPA-NE QAPP Worksheet #11b – EPA Rev. 10/99
Measurement Performance Criteria Table

| Sampling Procedure | QC Sample and/or Activity Used to Assess Measurement Performance | Measurement Performance Criteria | Data Quality Indicators (DQIs) | Analytical Method/SOP |
|--|---|--|--------------------------------|-----------------------|
| Total Phosphorus and Nitrogen Analysis | Digestion (Method) Blank | Not greater than 4 µg P/L and 30 µg N/L | Bias | 016 |
| | Sample Replication | Not greater than 25%RPD (Replicate from different sample cups) Not greater than 25%RPD (Replicate digestions) | Precision | |
| | Calibration | R ² of calibration linear regression not less than 0.990 | Accuracy | |
| | Matrix Spike | 80 – 120 % recovery | Bias | |
| | EPA Water Pollution Proficiency Test Study (Analysis of Unknowns) | 2 standard deviation | Accuracy/ Comparability | |
| | Laboratory Control Samples (Purchased External Standards) | Not greater than 20%D | Accuracy/ Comparability | |
| | Standards as Samples (Calibration check) | Not greater than 20%D | Accuracy | |
| | Check Temperature of Water Bath | At least 100 °C (boiling) for 15 minutes | Bias | |

2.6.3 Fixed Laboratory Analytical Method/SOP Reference Table – Worksheet 20

| EPA-NE QAPP Worksheet #20 – EPA Rev. 10/99 Fixed Laboratory Analytical Method/SOP Reference Table | | | |
|--|--|-----------------------------|--|
| Reference Number (SOP Number) | Title, Revision Date and/or Number | Analytical Parameter | Instrument |
| 024 | Fecal coliform Analysis Using Colilert-18 IDEXX Method Rev. 2: 01/16 | Fecal coliforms | Incubator – 44.5 °C ThermoScientific Heratherm incubator Autoclave UV Sterilization box |
| 018 | Enterococci Analysis Using Enterolert IDEXX Method, Rev. 3: 01/16 | Enterococci | Incubator – 41 °C ThermoScientific Heratherm incubator UV Indicator lamp Autoclave UV Sterilization box |
| 020 | Heterotrophic Plate Count for Quanti-tray IDEXX Method, Rev. 3: 5/20 | Heterotrophic plate count | Incubator – 35 °C Thermolyne Type 142300 incubator UV Indicator Lamp Autoclave UV Sterilization box |
| 010 | Alkalinity and pH Procedures, Rev. 4: 11/16 | Alkalinity and pH | Hanna Instruments Model HI 902 automatic potentiometric titrator |
| 017b | Salinity Analysis Using a Digital Refractometer, Rev. 1: 01/23 | Salinity | Hanna Instruments HI96822 Seawater Refractometer |

| EPA-NE QAPP Worksheet #20 – EPA Rev. 10/99 Fixed Laboratory Analytical Method/SOP Reference Table | | | |
|--|---|--|---|
| Reference Number (SOP Number) | Title, Revision Date and/or Number | Analytical Parameter | Instrument |
| 012 | Chlorophyll-a Analysis, Welschmeyer Method, Rev. 6: 04/18 | Chlorophyll-a | Fluorometer – Turner Designs Trilogy® |
| 013 | Chloride Analysis, Rev. 5: 11/2016 | Chloride (Cl ⁻) | Astoria®-Analyzer Model 303A Segmented Continuous Flow Autoanalyzer |
| 014 | Ammonia Analysis, Rev. 5: 11/2016 | Ammonia (NH ₃ -N) | Astoria®-Analyzer Model 303A Segmented Continuous Flow Autoanalyzer |
| 015 | Orthophosphate and Nitrate + Nitrite Analysis, Rev. 5: 12/16 | Orthophosphate (PO ₄ -P) and Nitrate + Nitrite (NO ₃ +NO ₂ -N) | Astoria®-Analyzer Model 303A Segmented Continuous Flow Autoanalyzer |
| 016 | Total Phosphorus and Nitrogen Analysis, Rev. 5: 12/16 | Total phosphorus (P) and total nitrogen (N) | Astoria®-Analyzer Model 303A Segmented Continuous Flow Autoanalyzer Water baths (Labline and Precision Scientific 83) |

Notes:

No SOP was modified for project work as this QAPP is for general laboratory procedures and not associated with a specific project.

All Fixed laboratory analytical methods are for definitive data

All fixed laboratory analytical methods are performed by URIWW laboratory

No analytical methods have Region 1 NESTS Method Codes.

2.6.4 Fixed Laboratory Instrument Maintenance and Calibration Table - Worksheet 21

| EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99 Fixed Laboratory Instrument Maintenance and Calibration Table | | | | | | | |
|---|--|--|----------------|--|---|---------------------------|---------------|
| | | Maintenance, Testing and Inspection Activities | | | | | |
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| Heterotrophic plate count | Incubator – 35 °C Thermolyne 142300 incubator | Check temperature | Each time used | 35 +/- 0.5 °C | Adjust temperature control | E. Herron | 018, 020, 024 |
| Fecal coliform, Heterotrophic plate count and Enterococci | UV sterilization box | Check UV sterilizer light efficiency using UV light meter | Quarterly | Not less than 70% of initial efficiency | Replace UV lamps | E. Herron | 018, 020, 024 |
| Fecal coliform, Heterotrophic plate count and Enterococci | Autoclave | Check temperature and pressure Use sterile indicator strips | Each time used | Must reach set temperature Must indicate “OK” | Contact professional to provide maintenance service | E. Herron | 018, 020, 024 |
| Fecal coliform, Heterotrophic plate count and Enterococci | Autoclave | Confirm sterilization using spore strips | Monthly | No growth | Contact professional to provide maintenance service | E. Herron | 018, 020, 024 |

EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table

| | | Maintenance, Testing and Inspection Activities | | | | | |
|--|--|--|----------------------|---------------------|--|---------------------------|------------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| Fecal coliforms, | Incubator – 44.5 °C ThermoScientific Heratherm incubator | Check temperature | Each time used | 44.5 +/- 0.2 °C | Adjust temperature control | E. Herron | 024 |
| Enterococci | ThermoScientific Heratherm incubator | Check temperature | Each time used | 41 °C +/- 0.5 °C | Adjust temperature control | E. Herron | 018 |
| Fecal coliform, Heterotrophic plate count and Enterococci | Flammables refrigerator | Check temperature | Daily when in use | 4 °C +/- 2°C | Adjust temperature | E. Herron | 018, 020, 024 |
| Fecal coliform, Heterotrophic plate count and Enterococci | IDEXX Quanti- Tray Sealer Model 2x | Make sure sealer seals each tray | Each time used | Tray is sealed | Clean according to SOP and if still not working, send for repair | E. Herron | 018, 024 |
| Heterotrophic plate count and Enterococci | UV light box (for counting cells) | Check that bulbs function | Daily when in use | Light turns on | Replace bulbs | E. Herron | 018, 020 |

**EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table**

| | | Maintenance, Testing and Inspection Activities | | | | | |
|-------------------|---|--|---|--|--|---------------------------|--------------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| Alkalinity and pH | pH & alkalinity Meter – Hanna Instruments HI 902 | Calibrate | Each time used | Electrode Efficiency >96% | Replace standards then if calibration still a problem replace the electrode filling solution, then replace electrode | URIWW Staff | 010 |
| | pH electrode – Hanna HI1053B | Refill electrode with saturated KCl solution, HI70300S | Check before each use | KCl solution is within ¼ inch of top of electrode and filling hole is open | Re-fill electrode as needed | URIWW Staff | 010 |
| Salinity | Hanna Instruments HI 96822 Seawater Refractometer | Check that prism is not damaged: compare with 0 and 20 ppt standards | Each time used | Prism is clean and not scratched or cracked | Send instrument to manufacturer for repair, calibrate & use another refractometer | E. Herron | 017b |
| Chlorophyll-a | Fluorometer – Turner Designs Trilogy® | Calibrate | Calibrate yearly, Check calibration daily | Daily – Not greater than 15%D | Re-calibrate and then replace light source if calibration continues to drift | URIWW Staff E. Herron | 012 |
| | Flammables freezer | Check temperature | Daily when in use | -20 °C +/- 5°C | Adjust temperature | E. Herron | 012, 018, 020, 024 |

EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table

| | | Maintenance, Testing and Inspection Activities | | | | | |
|--|---|--|----------------|--|--|---------------------------|--------------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| Chloride, Ammonia, Orthophosphate and Nitrate + Nitrite, Total Phosphorus and Total Nitrogen | Astoria®-Analyzer Model 303A Segmented Continuous Flow Autoanalyzer | Calibrate | Each time used | R ² of calibration linear regression not less than 0.99 | Re-calibrate | K. Addy | 013, 014, 015, 016 |
| | | Check analytical tubing | Each time used | No cracks or clogs | Replace affected tubing | K. Addy | |
| | | Check reagents flows | Each time used | No clogs in tubing causing pulsating flow | Replace affected tubing | K. Addy | |
| | | Check light source voltage | Each time used | < 70V and greater than reference voltage | Replace light source | K. Addy | |
| | | Check baseline | Each time used | Should be smooth | Replace tubing/trouble shoot instrument using instruction manual | K. Addy | |
| | | Check inter sample bubble shape | Each time used | Bubble shape is uniform | Adjust tubing, flow or reagents | K. Addy | |
| | | Check peak height and shape | Each time used | Check that peaks are not off scale | Dilute samples | K. Addy | |

**EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table**

| | | Maintenance, Testing and Inspection Activities | | | | | |
|--|---|--|-------------------|--|------------------------|---------------------------|--------------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| Chloride, Ammonia, Orthophosphate and Nitrate + Nitrite, Total Phosphorus and Nitrogen | Indesit refrigerator (Rm 018) | Check temperature | Daily when in use | 4 °C +/- 2°C | Adjust temperature | E. Herron | 013, 014, 015, 016 |
| | Small brown refrigerator (Rm 018) | Check temperature | Daily when in use | 4 °C +/- 2°C | Adjust temperature | E. Herron | |
| Total Phosphorus and Total Nitrogen | Precision Scientific Company Water Bath | Check Temperature | Each time used | At least 100 °C (boiling) for 15 minutes | Adjust temperature | E. Herron | 016 |
| | Labline Water Bath | Check Temperature | Each time used | At least 100 °C (boiling) for 15 minutes | Adjust temperature | E. Herron | 016 |

**EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table**

| | | Maintenance, Testing and Inspection Activities | | | | | |
|------------|--|---|-------------------|---|---|---------------------------|---------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| All assays | Walk-in Refrigerator (Hartford Duracool) | Check temperature | Daily when in use | 4 °C +/- 2°C | Report to Department Admin. Assist. To schedule repair | E. Herron | NA |
| | Aries Vaponics Water Filtration Units | Check resistively Check for leaking | Daily when in use | At least 17 Megaohms | Replace cartridges, contact American Aqua Systems for professional repair | E. Herron | 002 |
| | Laboratory Reagent Grade Water | Check concentration of lead, cadmium, chromium, copper, nickel and zinc | Annually | Metals individually at a level of less than 0.05 mg/L and the sum of all metals concentrations less than 0.1 mg/L | Replace water filtration cartridges or upgrade water treatment | E. Herron | 002 |
| | Laboratory Reagent Grade Water | Heterotrophic plate count | Monthly | Less than 500 CFU/mL | Replace water filtration cartridges or upgrade water treatment | E. Herron | 020 |
| | Frigidaire Refrigerator (Rm. 002) | Check temperature | Daily when in use | 4 °C +/- 2°C | Adjust temperature | E. Herron | |

EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99
Fixed Laboratory Instrument Maintenance and Calibration Table

| | | Maintenance, Testing and Inspection Activities | | | | | |
|---------------------------|--|---|---|---|--|---------------------------|---------------|
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| All assays (Continued) | Frigidaire Freezer (Rm 002) | Check temperature | Daily when in use | -20 °C +/- 5°C | Adjust temperature | E. Herron | |
| | Laboratory Thermometers | Check calibration of thermometer against NIST calibrated reference thermometer over the range the thermometer is generally utilized over or the temperature of interest. | Annually – glass thermometers and quarterly – all other thermometers | If checking calibration over a range, correction factors should be constant. | Any correction factors will be noted on the thermometer, if thermometer error in non- linear thermometer will be disposed of. | Laboratory staff | |
| | NIST calibrated reference thermometers | Re-calibration | Every 5 years | Determined by professional | Update correction factors or obtain new reference thermometer | E. Herron | |
| | Analytical Balance – Mettler Toledo AB 104 | Check calibration | Each day used | Not greater than 10%D for 50 mg weight and 1% for 20 g weight | Contact professional to provide maintenance and calibration service | E. Herron | 019 |
| | Analytical Balance – Mettler Toledo AB 104 | Check tare | Each Day used | +/-0.0001 g | Level, tare and check | E. Herron | 019 |

| EPA-NE QAPP Worksheet #21 – EPA Rev. 10/99 Fixed Laboratory Instrument Maintenance and Calibration Table | | | | | | | |
|---|--|--|--------------------|-------------------------------|---|---------------------------|---------------|
| | | Maintenance, Testing and Inspection Activities | | | | | |
| Activity | Instrument | Activity | Frequency | Acceptance Criteria | Corrective Action (CA) | Person Responsible for CA | SOP Reference |
| All assays (Continued) | Analytical Balance – Mettler Toledo AB 104 | Re-calibration | Every 3-5 years | Determined by professional | Re-calibration of balance or service balance if necessary | E. Herron | 019 |

Note: Maintenance and service record forms are in Appendix D.

Fixed Laboratory Analytical QC Sample Table –Worksheet #24 a

| Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99 | | | | | |
|--|---|---|---|------------------------------|------------------------------|
| FECAL COLIFORMS | | | | | |
| Medium/Matrix | Water | Analytical Method/ SOP Reference | | SOP 024 | |
| Sampling SOP | NA | Laboratory Name | | URIWW | |
| Concentration Range (without dilution) | <1 to 2,419 CFU/100 mL using a Quanti-Tray/2000 <1 to 200 CFU/100 mL using a Quanti-Tray | No. of Sample Locations | | NA | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Method Blank | 2/run or 2/100 samples, whichever is greater | Less than 1 cell/100 mL | Samples re-analyzed, data qualified as outside holding time | E. Herron | Bias |
| Sample Replication | 25% of non-diluted samples | 20%RPD | Data qualified | E. Herron | Precision |
| | Diluted samples replicated by comparing samples at different dilutions | 20%RPD | Data qualified | E. Herron | Precision |
| Sample Bottle Sterility Check | 1 per batch of bottles | Less than 1 cell/100 mL | Re-sterilize bottles | E. Herron | Bias/False positives |
| UV sterilizer light efficiency | Biennially | Less than 1 cell/100 mL | Replace UV lights | E. Herron | False Positives |
| Mixing bottle sterility check | 1 per batch of bottles | Less than 1 cell/100 mL | Re-sterilize bottles | E. Herron | Bias/False positives |
| Positive Trays | 2/run | Positive growth | Samples re-analyzed, data qualified as outside holding time | E. Herron | False Negatives/Bias |
| QA Checks of new reagent lots | 1 set of QC per new lot | Positive growth for positive, no growth for negatives | Return reagent batch if reference lots exhibits correct growth but new lot does not. Re-run if reference doesn't exhibit growth | E. Herron | False negatives |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
ENTEROCOCCI

| Medium/Matrix | Water | | Analytical Method/ SOP Reference | SOP 018 | |
|--|---|---------------------------------------|--|---------------------------------|---------------------------------|
| Sampling SOP | NA | | Laboratory Name | URIWW | |
| Concentration Range (without dilution) | <1 to 2,419 CFU/100 mL using a Quanti-Tray/2000 <1 to 200 CFU/100 mL using a Quanti-Tray | | No. of Sample Locations | NA | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Method Blank | 2/run | Less than 1 cell/100 mL | Samples re-analyzed, data qualified as outside holding time | E. Herron | Bias |
| Sample Replication | 1 sample in 5 (20% of samples) | 20%RPD | Data qualified | E. Herron | Precision |
| Sample Bottle Sterility Check | 1 per batch of bottles | Less than 1 cell/100 mL | Re-sterilize bottles | E. Herron | Bias/False positives |
| UV sterilizer light efficiency | Biennially | Less than 1 cell/100 mL | Replace UV lights | E. Herron | False Positives |
| Positive Trays | 1 per run | Positive growth | Samples re-analyzed, data qualified as outside holding time | E. Herron | False Negatives/Bias |
| Mixing bottle sterility check | 1 per batch of bottles | Less than 1 cell/100 mL | Re-sterilize bottles | E. Herron | Bias/False positives |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note:
No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
Heterotrophic Plate Count

| Medium/Matrix | Water | Analytical Method/ SOP Reference | SOP 026 | | |
|--|------------------------|---------------------------------------|------------------------|---------------------------------|---------------------------------|
| Sampling SOP | NA | Laboratory Name | URIWW | | |
| Concentration Range (without dilution) | <1 to 200 colonies/ mL | No. of Sample Locations | NA | | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| UV sterilizer light efficiency | Biennially | Less than 1 cell/100 mL | Replace UV lights | E. Herron | False Positives |
| Mixing bottle sterility check | 1 per batch of bottles | Less than 1 cell/100 mL | Re-sterilize bottles | E. Herron | Bias/False positives |

Note:
No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99

ALKALINITY AND pH

| Medium/Matrix | Water | | | Analytical Method/ SOP Reference | SOP 010 |
|--|--|--|--|----------------------------------|---------------------------------|
| Sampling SOP | NA | | | Laboratory Name | URIWW |
| Concentration Range (without dilution) | pH ambient & marine samples: 3 – 12 SU | Alkalinity Ambient water : <0.1 – 30 mg/L CaCO ₃ | No. of Sample Locations | | NA |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Calibrate pH meter | Each time used | Electrode Efficiency >96% | Replace standards, change electrode filling solution, replace the electrode | URIWW Staff | Accuracy |
| Standards as Samples (check of calibration using pH 7.0 standard) | At beginning and end of each run | Change in standard not greater than +/- 0.2 SU | Re-check calibrant (7.0 pH buffer), then recalibrate and re-analyze affected samples if necessary | URIWW Staff | Accuracy/Precision |
| Laboratory Control Standard (LCS) for Alkalinity | At beginning of each run | Mean of LCS is 85-115% of true value stock LCS | Make new diluted (20 ppm) LCS, purchase new LCS, re-analyze | URIWW Staff | Accuracy/Precision |
| Sample Replication | 10% | pH – not greater than +/- 0.5 S.U., Alkalinity – not greater than 25%D | Re-analyze sample, then recalibrate instrument and re-analyze affected samples if necessary | URIWW Staff | Precision |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown for pH and alkalinity | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
SALINITY

| Medium/Matrix | Water | Analytical Method/ SOP Reference | SOP 017b-Salinity Analysis Using a Digital Refractometer | | |
|---|------------------------------|----------------------------------|--|------------------------------|------------------------------|
| Sampling SOP | NA | Laboratory Name | URIWW | | |
| Concentration Level (undiluted samples) | Marine samples: 0.4 - 40 ppt | No. of Sample Locations | NA | | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Sample Replication | 100% | Not greater than 2 ppt different | Repeat refractometer measurement a third time. If still greater than 2 ppt different note deviation on project data sheet | URIWW Staff | Precision |
| Sample Comparison | 50% | Not greater than 2 ppt different | Re-analyze sample by refractometer. If difference still greater than 2 ppt then recalibrate refractometer and measure DI and 20 ppt standards. If still greater than 2 ppt different it will be assumed the refractometer is in error and it will be replaced. | URIWW Staff | Accuracy/ Comparability |

Note:
No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

**Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
CHLOROPHYLL-a**

| Medium/Matrix | Water | Analytical Method/ SOP Reference | SOP 012 | | |
|---|---|---|---|---------------------------------|---------------------------------|
| Sampling SOP | NA | Laboratory Name | URIWW | | |
| Concentration Level (undiluted samples) | Ambient and marine samples: <0.2 – 100 µg/L chlorophyll-a | No. of Sample Locations | NA | | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Calibrate fluorometer | Calibrate yearly with purchased liquid chl standards, also record solid standard readings | N/A | track solid standard LCS values to note drift; Purchase additional liquid standard to compare against | URIWW Staff/ E. Herron | Accuracy |
| LCS (Check standard using solid standard) | 1/rack (38 samples) | Not greater than 15%D | Re-analyze on fluorometer, check value of primary standard, recalibrate if necessary, re-analyze associated samples | URIWW Staff | Accuracy |
| Method Blank | 1/rack (38 samples) | Not greater than 0.03 µg/L chlorophyll-a as read on the fluorometer | Re-analyze on fluorometer, then qualify samples associated with blank if necessary | URIWW Staff | Bias |
| Filter Blank | 1/rack (38 samples) | Not greater than 0.03 µg/L chlorophyll-a as read on the fluorometer | Re-analyze on fluorometer, then qualify samples associated with blank if necessary | URIWW Staff | Bias |

Note:
No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

**Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
CHLORIDE**

| | | | |
|---|--|----------------------------------|---------|
| Medium/Matrix | Water | Analytical Method/ SOP Reference | SOP 013 |
| Sampling SOP | NA | Laboratory Name | URIWW |
| Concentration Range (without dilution) | Ambient samples: 5 – 50 mg/L Cl ⁻ | No. of Sample Locations | NA |

| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
|--|--|--|--|------------------------------------|------------------------------------|
| Calibrate | Each time used Analyze a set of calibrants at start and end of run | R ² of calibration linear regression not less than 0.990 | Re-calibrate/start a new analytical run | K. Addy | Accuracy |
| Method Blank | 1 per 15 sample cups | Not greater than 2 mg/L Cl ⁻ | Re-analyze, then re-calibrate and re- analyze associated samples if necessary | K. Addy | Bias |
| Sample Replication | Samples aliquot from same cup – 100% Ambient samples poured into two separate cups – 10% of samples | Not greater than 20%RPD Not greater than 20%RPD | Re-analyze samples, if still greater than QC objective then note deviation on project data sheet | K. Addy | Precision |
| Matrix Spike | ~1 per 45 sample cups | 80 – 120% recovery | Re-analyze, if still not acceptable but all other spikes and calibration checks are acceptable, flag data. If other spikes unacceptable spike another sample to determine if matrix is interfering, then consult laboratory manager. | K. Addy | Bias |
| LCS (Purchased External Standards) | 1 or 2 per rack of 90 sample cups | Not greater than 20%D | Re-analyze standard, then recalibrate instrument and re-analyze associated samples if still greater than 20%D | K. Addy | Accuracy/ Comparability |
| Standards as Samples (Calibration Check) | 1 per 15 sample cups | Not greater than 20%D | Re-analyze standard, then recalibrate instrument and re-analyze associated samples if still greater than 20%D | K. Addy | Accuracy |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

**Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
AMMONIA**

| | | | | | |
|--|---|---|--|------------------------------------|------------------------------------|
| Medium/Matrix | Water | Analytical Method/ SOP Reference | | SOP 014 | |
| Sampling SOP | NA | Laboratory Name | | URIWW | |
| Concentration range (without dilution) | Ambient and marine samples: <40 – 500 µg/L NH ₃ -N | | No. of Sample Locations | | NA |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Calibrate | Each time used. Analyze a set of calibrants at start and end of run | R ² of calibration linear regression not less than 0.990 | Re-calibrate/start a new analytical run | K. Addy | Accuracy |
| Method Blank | 1/15 samples | Not greater than 30 µg/L NH ₃ -N | Re-analyze, then re-calibrate and re-analyze associated samples if necessary | K. Addy | Bias |
| Sample Replication | Ambient & marine samples poured into two separate cups – 1-2 per 15 sample cups | Not greater than 20%RPD | Re-analyze samples, if still greater than QC objective then note deviation on project data sheet | K. Addy | Precision |
| LCS (Purchased External Standards) | 1 or 2 per rack of 90 sample cups | Not greater than 20%D | Re-analyze standard, if still outside QC objective recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy/ Comparability |
| Standards as Samples (Calibration Check) | 1/15 sample cups | Not greater than 20%D | Re-analyze standard, if still outside QC objective recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy |
| Matrix Spike | 1 per ~45 sample cups | 80 – 120% recovery | Re-analyze, if still not acceptable but all other spikes and calibration checks are acceptable, flag data. If other spikes unacceptable spike another sample to determine if matrix is interfering, then consult laboratory manager. | K. Addy | Bias |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | Yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
ORTHOPHOSPHATE AND NITRATE + NITRITE

| | | | | | |
|--|---|--|--|------------------------------------|------------------------------------|
| Medium/Matrix | Water | Analytical Method/ SOP Reference | | SOP 015 | |
| Sampling SOP | NA | Laboratory Name | | URIWW | |
| Concentration range (without dilution) | Nitrate/Nitrite: Ambient and marine samples: <15 – 1000 µg/L NO ₃ /NO ₂ -N Orthophosphate: Ambient and marine samples: <4– 100 µg/L PO ₄ -P | No. of Sample Locations | | NA | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Calibrate | Each time used, analyze a set of calibrants at beginning and end of run | R ² of calibration linear regression not less than 0.990 | Re-calibrate/start a new analytical run | K. Addy | Accuracy |
| Method Blank | 1 per 15 sample cups | Not greater than 2 µg/L PO ₄ -P and 20 µg/L NO ₃ /NO ₂ -N | Re-analyze, then re-calibrate and re-analyze associated samples if necessary | K. Addy | Bias |
| Sample Replication | Ambient & marine samples poured into two separate cups 1 or 2 per 15 sample cups | Not greater than 20%RPD | Re-analyze samples, if still greater than QC objective then note deviation on project data sheet | K. Addy | Precision |
| LCS (Purchased External Standards) | 1 or 2 per rack of 90 sample cups | Not greater than 20%D | Re-analyze standard, if still outside QC objective recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy/ Comparability |
| Standards as Samples (Calibration Check) | 1 per15 sample cups | Not greater than 20%D | Re-analyze standard, if still outside QC objective recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy |
| Matrix Spike | 1 per 45 sample cups | 80 – 120% recovery | Re-analyze, if still not acceptable but all other spikes and calibration checks are acceptable, flag data. If other spikes unacceptable spike another sample to determine if matrix is interfering, then consult laboratory manager. | K. Addy | Bias |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99
TOTAL PHOSPHORUS AND NITROGEN

| | | | | | |
|--|--|---|--|------------------------------------|------------------------------------|
| Medium/Matrix | Water | Analytical Method/ SOP Reference | | SOP 016 | |
| Sampling SOP | NA | Laboratory Name | | URIWW | |
| Concentration range (without dilution) | Total N Ambient and marine samples: <40 – 1000 µg/L N Total P Ambient and marine samples: <4 – 100 µg/L P | No. of Sample Locations | | NA | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| Calibrate | Each time used; analyze a set of calibrants at the beginning and end of run | R ² of calibration linear regression not less than 0.990 | Re-calibrate/start a new analytical run | K. Addy | Accuracy |
| LCS (Purchased External Standards) | 2-3 per rack of 90 sample cups | Not greater than 20%D | Re-analyze standard, then if still outside QC criteria recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy/ Comparability |
| Standards as Samples (Calibration Check) | 1 per 15 sample cups | Not greater than 20%D | Re-analyze standard, if still outside QC criteria recalibrate instrument and re-analyze associated samples | K. Addy | Accuracy |
| Matrix Spike | 1 per ~30 sample cups | 80 – 120% recovery | Re-analyze, if still not acceptable but all other spikes and calibration checks are acceptable, flag data. If other spikes unacceptable spike another sample to determine if matrix is interfering, then consult laboratory manager. | K. Addy | Bias |
| Method/Digestion Blank | ~1 per 30 sample cups | Not greater than 5 µg/L P and 30 µg/L N | Note on project data sheet. If all digestion blanks are outside acceptable range, run is considered contaminated and data marked accordingly. | K. Addy | Bias |
| Sample Replication | ~20% replicate digestions of marine and ambient samples | Not greater than 25%RPD | Re-analyze sample, if still outside QC criteria note deviation on project data sheet | K. Addy | Precision |

| Fixed Laboratory Analytical QC Sample Table – EPA NE QAPP Worksheet #24 a – EPA Rev. 10/99 TOTAL PHOSPHORUS AND NITROGEN | | | | | |
|---|--|------------------------------------|-----------------------------|------------------------------------|------------------------------------|
| Medium/Matrix | Water | Analytical Method/ SOP Reference | | SOP 016 | |
| Sampling SOP | NA | Laboratory Name | | URIWW | |
| Concentration range (without dilution) | Total N Ambient and marine samples: <40 – 1000 µg/L N Total P Ambient and marine samples: <4 – 100 µg/L P | No. of Sample Locations | | NA | |
| Laboratory QC: | Frequency/ Number | Method/SOP QC Acceptance Limits | Corrective Action (CA) | Person(s) Responsible for CA | Data Quality Indicator (DQI) |
| EPA Water Pollution Proficiency Test Study – Analysis of unknown | yearly | 2 standard deviation | Data reported to laboratory | NA | Accuracy/ Comparability |

Note: No measurement performance criteria are provided in this table as this QAPP is for general laboratory procedures and not associated with a specific project.

3.0 SAMPLE HANDLING, TRACKING AND CUSTODY REQUIREMENTS

Sample bottles are sent out prelabeled with each bottle identified by label color and / or bottle type. Labels include: Site name, sampling depth, date expected (exact sample date circled or written in), time sampled (line for writing in) and sample type (i.e. bacteria, unfiltered, filtered, etc.) to ensure that analyses are completed on appropriate types of sample. Bottles are brought to the laboratory in coolers with cold packs or ice. A chain-of-custody (COC) form or sample log sheet are completed for each set of samples by the person(s) responsible for collection and/or delivery of the samples to the laboratory. Sample log sheets are generally provided by the URIWW Laboratory. The COC form will include the following information:

1. Collection Period & type of sample (lake, river, salt ponds sites etc.)
2. Monitoring Location
3. Person(s) responsible for transporting samples
4. Date of sample collection, time of sample collection
5. Number and type of sample bottles
6. Temperature of bacteria sample upon receipt

A technician will be responsible for checking that the samples listed on the sample log sheet correspond correctly with the samples received. A copy of the sample log sheet will be maintained in the project file. The set of logs sheets for each sample collection period are scanned and pdfs kept in electronic files for that year. A portion of a sample log sheet is in Figure 3.

Figure 3. Portion of sample log sheet

| 2020 Lakes and Ponds Log Sheet: September Water Collection | | | | | | | | | | | | |
|---|-----------|-----------|------|----------|------|------------|------------|----------|---------|-------|---------|------------|
| You should have all of the bottles/bags listed with your monitoring location. | | | | | | | | | | | | |
| Monitoring Location | Your name | Date | Time | Date | Time | Plastic | Brown | Plastic | Plastic | Chl-a | Trib | Temp C |
| | | Collected | | Received | | Unfiltered | Unfiltered | Filtered | Sterile | baggy | Bottles | at receipt |
| Annaquatucket Mill P | | | | | | 1 | 1 | 1 | 1 | | | |
| Beach Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Belleville P - Lower | | | | | | 1 | 1 | 1 | 1 | | | |
| Carr Pond (WG) | | | | | | 2 | 2 | 1 | 1 | | | |
| Flat River Reservoir | | | | | | 2 | 2 | 1 | 1 | | | |
| Georgiaville Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Long Pond (SK) | | | | | | 2 | 2 | 1 | 1 | | | |
| Mashapaug Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Mishnock Lake | | | | | | 2 | 2 | 1 | 1 | | | |
| Pascoag Reservoir | | | | | | 2 | 2 | 1 | 1 | | | |
| Pasquisset Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Posnegansett Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Posnegansett Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| Prince's Pond | | | | | | 2 | 2 | 1 | 1 | | | |
| RWP #1 -Roosevelt Lake inflow | | | | | | 1 | 1 | 1 | 1 | | | |

3.1 Laboratory Sample Tracking

Once samples are accepted into the laboratory, internal laboratory tracking is accomplished by placing colored adhesive dots onto each sample bottle to designate the assays to be completed (Figure 4). After each process and/or analysis has been completed for a sample, the colored dot is removed. Unfiltered and filtered samples may be stored until after nutrient analyses are completed and reviewed.

Figure 4 – Internal Laboratory Tracking Scheme

| General type of analysis | Type of bottle | Sample bottle label information | Colored label dot (Avery Cat. #) |
|--|---------------------------------------|--|--|
| Bacteria, shallow, unfiltered sample collected directly into sample bottle | HDPE autoclaved 250 ml | White colored label: Site name Date expected Time of collection ____ BACTERIA | No dots STERILE label affixed across lid before use |
| Bacteria, shallow, unfiltered sample collected directly into sample bottle | HDPE autoclaved 250 ml | Blue colored label: Trib @ Site name Date expected Time of collection ____ BACTERIA | No dots STERILE label affixed across lid before use |
| pH and Alkalinity, shallow, unfiltered water sample from lake sites | HDPE 500 ml | Magenta colored label Site name 1m (or mid-depth) Date expected Time of collection ____ pH and Alkalinity | Half a yellow $\frac{3}{4}$ " dot (UNV401141ND) |
| pH, shallow, unfiltered water sample from any sites | 100 ml "specimen" cup | Magenta colored label Site name 1m (or mid-depth) Date expected Time of collection ____ pH - UNFILTERED | Quarter of a yellow $\frac{3}{4}$ " dot (UNV401141ND) |
| pH, deep (>1m), unfiltered water sample from any sites | 100 ml "specimen" cup | Green colored label Site name 1m (or mid-depth) Date expected Time of collection ____ UNFILTERED | Quarter of a yellow $\frac{3}{4}$ " dot (UNV401141ND) |
| pH, shallow, unfiltered water sample from tributary stream or cove to lake sites | 100 ml "specimen" cup | Blue colored label Trib @ Site name Date expected Time of collection ____ pH - UNFILTERED | Quarter of a yellow $\frac{3}{4}$ " dot (UNV401141ND) |
| Nutrients, shallow, unfiltered water sample from any sites for total phosphorus and total nitrogen | 125 – 250 ml brown glass, acid washed | Magenta colored label Site name 1m (or mid-depth) Date expected Time of collection ____ UNFILTERED | Neon green (AVE05052) If filtered water is needed, Light blue (AVE05050) with "F" written on it |

(continued)

Figure 4 – Internal Laboratory Tracking Scheme (continued)

| General type of analysis | Type of bottle | Sample bottle label information | Colored label dot (Avery Cat. #) |
|---|---------------------------------------|--|---|
| Nutrients, deep (>1m), unfiltered water sample from any sites for total phosphorus and total nitrogen | 125 – 250 ml brown glass, acid washed | Green colored label Site name 1m (or mid-depth) Date expected Time of collection ____ UNFILTERED | Neon green (AVE05052) and Light blue (AVE05050) with "F" written on it |
| Nutrients, shallow, unfiltered water sample from tributary stream or cove to lake sites total phosphorus and total nitrogen | 125 – 250 ml brown glass, acid washed | Blue colored label Trib @ Site name Date expected Time of collection ____ UNFILTERED | Neon green (AVE05052) and Light blue (AVE05050) with "F" written on it |
| Nutrients, filtered, water sample from any site and depth | 125 ml HDPE, acid washed | Yellow colored label Site name 1m, mid-depth, or ____m DEEP Date expected Time of collection ____ FILTERED | Neon orange (AVE05062) – nitrate+nitrite N; Light blue (AVE05050) – Dissolved P; White quartered (AVE05408) – ammonia; Neon red (AVE05051) - chloride |

3.2 Acceptance of Expendable Laboratory supplies

All expendable laboratory supplies such as chemical reagents and sample bottles will be inspected upon arrival by either Elizabeth Herron or Kelly Addy. Packages containing damaged material or packages that were open upon arrival will not be accepted. Chemicals will be marked with the date of acceptance as well as the date they are opened.

3.3 Sample Handling System – Worksheet 16

| |
|--|
| EPA-NE QAPP Worksheet #16 – EPA Rev. 10/99 Sample Handling System |
| SAMPLE COLLECTION, PACKAGING AND SHIPMENT |
| <p>Sample Collection: Various persons, staff or trained volunteers</p> <p>Sample Packing: Person(s) responsible for sample collection, samples kept in a cooler with ice packs</p> <p>Coordination of Delivery: Person(s) responsible for sample collection, delivered same day or next day with prior URIWW approval</p> <p>Type of Shipment: Driven to URI Watershed Watch laboratory or to a designated collection point for transport to URI</p> |
| SAMPLE RECEIPT AND ANALYSIS |
| <p>Responsible Organization: University of Rhode Island Watershed Watch Laboratory (URIWW)</p> <p>Sample Receipt: URIWW Staff</p> <p>Sample Custody and Storage: URIWW Staff</p> <p>Sample Preparation: URIWW Staff</p> <p>Sample Determinative Analysis: URIWW Staff</p> |
| SAMPLE ARCHIVAL |
| <p>Field Sample Storage (No. of days from sample collection): Dependent upon analysis – Refer to analyte-specific SOPs (Appendix A)</p> <p>Sample Extract/Digestate Storage: for TP & TN analysis <24 hours after digestion. Filtered samples frozen within hours, thawed within 16 hours of analysis. See also analyte-specific SOPs (Appendix A)</p> |
| SAMPLE DISPOSAL |
| <p>Responsible Organization and personnel: URIWW / URIWW Staff upon approval of E. Herron or K. Addy</p> |

4.0 PROJECT DOCUMENTATION AND RECORDS

All sample log sheets will be retained by the URI Watershed laboratory in laboratory file cabinets. Log sheets are also scanned and saved as pdfs in appropriate folders on URI servers. All hard copy sample data sheets and sample preparation worksheets are as discussed in each analyte-specific SOP under Section 7.0. Additional documentation will also be retained in project files as needed. For assays that produce electronic files, the electronic file will be stored on URI servers. Summary files will be produced and stored in both electronic and hard copy forms.

Project files are maintained in the main URIWW laboratory in the URI Coastal Institute by Elizabeth Herron and Kelly Addy. This location is locked when staff is not present. Electronic data are stored on shared drive on a University of Rhode Island server, with limited access. The server is backed up daily according to institution practices. All laboratory data (electronic and hard copy) are retained for at least 20 years.

Data will also be maintained online at the URI Watershed Watch website <https://web.uri.edu/watershedwatch/data/> in a variety of format for the use by researchers and the public. A combined datafile (1988 – 2021 currently) is available as a csv file (see <https://web.uri.edu/watershedwatch/data/historic-data/csv-data-files/>). To facilitate downloading of specific year data, ArcGIS or R code tools will be used to create annual files from that combined file. Data will be presented using a variety of dashboard tools as well.

5.0 DATA VALIDATION

No general quality management reports are prepared. Laboratory data management follows SOP 027 (Appendix A) and is summarized here. During the analysis of samples, the technician completing sample analysis is responsible for recording any problems with meeting measurement performance criteria (Section 2.6.2) and/or instrument operational issues. Reruns of analyses that do not meet measurement criteria are completed when possible. If the rerun value meets criteria, that value is used. If rerun values are not possible, any failure of a sample to meet defined measurement performance criteria should be recorded and the data flagged for further review upon data entry and final data validation. Any analyses that do not meet standard operating procedures, for example being run after designated hold time, will be recorded and included in the comment field for the data entry.

Technicians may also proactively address possible analysis errors for some parameters. For example, chlorophyll analyses usually include four (4) filters for a particular sample date for a site. If the measured value for a one (1) of those filters is significantly different from the other three (3) filters, it may be rerun to confirm sample value. The rerun value is recorded in the comments section of the data sheet to allow for that information to be tagged with the value when the data is entered. Other parameters with replicate analyses (i.e. pH, alkalinity, nutrients) permit the opportunity for confirming sample values when the differences between replicate values are greater than anticipated. In addition, those parameters allow for reruns when the results are significantly different than usual for the site. For that reason, data summary sheets from the previous season are posted near the Hanna meter (pH and alkalinity) and the Astoria Pacific rapid flow analyzer to allow for comparison during sample runs. Annual summary sheets are also available in the microbiology lab to ensure that the appropriate sample volumes /

Quanti-trays are used to get results within the acceptable counting range, minimizing values having to be reported as greater than.

Data generated by each analysis is internally validated by either Ms. Herron or Ms. Addy. The data validation process occurs at each step from data production to posting. When the data are keyed into Microsoft Excel files missing results or values outside of the anticipated range are identified and reported to allow for reruns or to find missing samples if possible. After data has been entered into the appropriate file, URIWW staff completes an initial check to be sure all data was entered correctly (ie. as written on the work sheet). This proofing process ensures that the data entered into the Excel file are as written on the worksheet. Outliers and inconsistencies are flagged for further review with corrections or verification completed by Ms. Herron.

Nutrient analyses undergo an additional validation step. Values for total phosphorus (P) are compared to orthophosphate-P (dissolved phosphorus). If dissolved P is greater than total P samples are reanalyzed for both constituents. A similar check is done to compare total nitrogen with the sum of nitrate/nitrite and ammonia. If the sum of dissolved N constituents is greater than total nitrogen then samples are re-analyzed.

Data may be compared to value obtained for similar samples analyzed in the past. Data may be tagged with a comment or other qualifier or may be discarded. The decision to discard data will be made by either Ms. Herron or Ms. Addy. Notations or qualifiers will be added to any results that are discarded.